

Service Manual for ECB counting scale

I . Phenomena of trouble

1) Trouble of part

The trouble phenomena are as following:

- a) Buzzer-----buzzer doesn't make sound, or sometimes make sound and sometimes doesn't make sound.
- b) LCD-----The display is faultily.
- c) Backlight-----The backlight doesn't lighten, or sometimes it lighten and sometimes doesn't lighten:
- d) Key -----The key don't work.

2) Trouble of load cell

The trouble phenomena are as following:

- a) The display data is drifting.
- b) The display data doesn't change, or the internal resolution value is not in the natural range.
- c) The internal resolution data drift badly.
- d) The initial resolution is drifting.

3) Trouble of power supply

The phenomenon is as following:

The scale doesn't work or the "low voltage" symbol comes out after turning on.

4) Trouble of the PCB


The phenomena are as following:

Trouble of main board


- There is nothing in the display when turn on the scale.
- The scale can not be powered on.
- The weighing is unstable.
- The internal resolution is out of the normal range.
- The display is abnormal, or shows nothing.
- The scale doesn't transmit any data.

II. To solve the trouble


1. The scale can not be powered on




2. No display after turn on:




3. RS232 don't transmit any data




4. Weighing is inaccurate or instable:



5. The keys don't work



6. The display is incomplete or exceptional.



Note: Since this series of balance is with high precision, it is very easy to be affected by climate and temperature. Sometimes the scale has the following phenomena:

1) Weighing is inaccurate.

The problem can be solved by re-calibration.

2) After self-check, the display shows error messages.

E1: Re-calibrate the scale to solve it.


E3: Re-calibrate the scale or replace the load cell to solve it.

3) If any one of main board and load cell is changed, the scale must be recalibrated.

III. Calibration procedure

- 1) Switch on, Press “83419” to enter into Calibration mode during self-checking. .

The windows show as follows:




Touch the pan by hand lightly to do the calibration if there is no need to change the weight units, accuracy and capacity.

- 2) Change and choose the weight units, accuracy and capacity by using the “ENTER” and the “MOVE” key.

The steps are as below:


Choose weight units:

- a) After the scale finish self- checking, press the “ENTER” key to enter into selecting weight units setting, and the display shows:




- b) Press the “MOVE” key to change the value : lb, 103g(=kg)
- c) Press the “ENTER” key to confirm and enter into next step (Choose accuracy value).

Note: If only the unit should be changed, then keeping press the “ENTER” key until the display shows as below, and press “CLEAR” key to exit calibration model.



Choose accuracy value:


- a) The display shows:




- b) Press the “MOVE” key to change the value. There are three system-preset (1, 2, 5) can be chosen.
c) Press the “ENTER” key to confirm and enter into next step (Choose capacity).

Choose capacity:


- a) The display shows:




- b) Use the 0-9 numerical keys to key in the capacity. The weighing unit is kg.
c) Press the “ENTER” key to confirm, at the same time enters into “Set the zero calibration value” mode, and the display shows:




- d) Use the numerical keys to key in the zero calibration value (0). Then Press the “ENTER” key to confirm and enter into “Set the first calibration value” mode, the display shows as below:




- e) Use the numerical keys to key in the first calibration value. Then Press the “ENTER” key to confirm and enter into “Set the second calibration value” mode, the display shows as below:



- f) Use the numerical keys to key in the first calibration value. Then Press the “ENTER” key to confirm and enter into “Set the third calibration value” mode, the display shows as below:



- g) Use the numerical keys to key in the first calibration value. Then Press the “ENTER” key to confirm and enter into the first step calibration mode. The display shows as below:



Calibration Mode:

- Touch the pan by hand lightly, several seconds later, the unit weight window shows “01”, and the weight window shows **the first calibration weight value** which should be put on for the first time. Put on the weight refer to the display.
- Several seconds later, the unit weight window shows “02”, and the weight window shows **the second calibration weight value** which should be put on for the second time. Put on the weight refer to the display.
- Several seconds later, the unit weight window shows “03”, and the weight window shows **the third calibration weight value** which should be put on for the third time. Put on the weight refer to the display.
- Several seconds later, the scale returns to counting mode. Then the calibration procedure is finished.
- Switch off, and take off all the weight, then switch on, the scale can be used.


Note; 1. There are four calibration weight values, CAP0-CAP3. You are supposed to set at least two calibration weight values.

2. If the number of calibration weight value is less than four, press SET key after set the last weight value and enter into the next calibration value setting mode. Then weight display window show “END”, press ENTER key to move to calibration mode.


3. The other way to enter into calibration mode is press CAL switch on the back of the indicator during the self-checking. By using this method, there is no need to change the unit, capacity and calibration weight.

3) Check the internal resolution

- a) Press “83419” or “CAL” switch during self-checking to enter into the calibration mode. The display shows as below;




- b) Press SET key to check the internal resolution. The display shows as below;



Current internal resolution


- c) Put on some weight. The weight display window shows the increased internal resolution, and the count window shows the total internal resolution which is equal to the original one and the increased internal resolution. The display shows as below;




Increased internal resolution

Current total internal resolution


- d) Press the SET key again, the display shows;



- e) Take away the load, the scale is back to calibration mode. The display shows;




After a while, you will hear a beeper. The display shows;



Simple calibration without weight

1. Turn on the scale, and key in "000419" during counting down (self-check) to zero to enter into Simple Calibration mode.

The displays will indicate as below eventually. (Take 150kg scale for instance)




Required weight (150 kg) for calibration

2. Keying in an error coefficient "R" (R= A/B, A=the correct value, B=the error value), then press the **ENTER** key, then the scale will be calibrated and return to normal counting mode.


- ★ Press **CLEAR** key to escape from calibration mode at any time.
- ★ The error coefficient is shown in "UNIT WEIGHT" window, and the decimal dot is fixed after the first digit, and it is hidden. So it is no need to key in the decimal dot.

For example:

- a) If the error coefficient is "0.89737", then the display will be as:



- b) If the error coefficient is "1.23456", then the display will be as:



Note: When the scale was calibrated in Country A, and it is used in Country B, the error coefficient (R) should be:

$$R = \text{gravity value in Country A} / \text{gravity value in Country B}$$